



# **TESTQUAL**

## **PROFICIENCY TESTING SCHEMES**

**TestQual, S.L.**

**(Proficiency Testing Schemes)**

*Pol. Industrial Oeste*

*Av. Principal, Parcela 21/1*

*CP 30169, San Ginés, Murcia*

*Telephone: 868 949 486 / 676 367 555*

**TestQual 211 PROTOCOL**

***Multiresidue pesticides in green beans***

# INDEX

0. GLOSARY AND ABBREVIATIONS.....	3
1. INTRODUCTION .....	3
2. OBJECTIVE .....	3
3. CALENDAR.....	3
4. REGISTER AND PARTICIPATION REQUEST (APPLICATION FORM).....	4
5. TEST MATERIAL. CONTROL AND DISTRIBUTION.....	6
6. CONCENTRATION RANGES, SIGMA OBJECTIVE AND ANALYTES.....	8
7. RESULTS EXPRESSION.....	11
8. STATISTICAL EVALUATION .....	12
9. EVALUATION REPORT/S .....	16
10. CONTACT .....	17
11. BIBLIOGRAPHY AND REFERENCES OF INTEREST .....	17

## 0. GLOSARY AND ABREVIATIONS

Text	Abbreviation
TestQual	TQ
Proficiency test	PT / P.T.
Limit Of Quantification	LOQ
Not Analysed	NA

## 1. INTRODUCTION

This document describes the **protocol** of the **TestQual 211** Proficiency Test (P.T.) in **green beans**, for the analysis of the **participant's choice** from the following group of **analytes**:

Analyte-group code	Grouped analytes
A	Multiresidue pesticides

In the present document is detailed how to start working with **TestQual**, send your application to participate, the statistic that will be applied and information about the evaluation report.

**TestQual, S.L.** is committed to maintaining confidentiality with the information of each laboratory from the beginning of the proficiency test.

## 2. OBJECTIVE

The objective of the **TestQual 211** Proficiency Test is to evaluate the quality and accuracy of the results sent by the participating laboratories. Because of this, proficiency testing is an essential element of laboratory quality assurance. It will help to control and detect errors in their results or methods of analysis.

## 3. CALENDAR

The following table shows the program for this proficiency test:

Date (YYYY/mm/dd)	Activity	Carried out by
2025/03/31	Deadline to reserve PT sample.	Participants
2025/05/05	Deadline to request participation*	Participants
2025/05/12-14	Sample delivery	<b>TestQual</b>
2025/06/02	Final date to submit results	Participants
2025/06/30	Evaluation/global report issued	<b>TestQual</b>

\*Acceptance of the participation between the deadline to reserve the PT sample and the deadline to request the participation will depend on availability of the proficiency test.

Participants are requested in the application to submit their LOQ/LOQs. For PTs with multiple groups of analytes, for each group that the participant has submitted LOQ/LOQs, if a certain percentage (described in our internal procedure) of the present/planned analytes is analysed, then the participation is accepted, and a laboratory code is generated and sent by email to confirm the participation acceptance to the user of the **TestQual**'s account.

The dates of this calendar and the definitive shipping distribution might be changed depending on the development of the proficiency test. The definitive distribution date is confirmed when the sample instructions are sent, which, besides the instructions for the sample, contains the deadline to submit the results (which might update and override the here displayed deadline).

The sample instructions are exclusively sent to inscribed participants. These possible changes would be notified to all participants through our website [www.TestQual.com](http://www.TestQual.com) and/or by email.

The **coordinator** of this proficiency test will be María Dolores García Valiente. Any question regarding the development of the proficiency test, the application status or any other query can be consulted by email to [testqual@testqual.com](mailto:testqual@testqual.com).

#### **4. REGISTER AND PARTICIPATION REQUEST (APPLICATION FORM)**

##### **NEW CLIENT**

If your laboratory has not participated before in one of our proficiencies tests you will have to register using the [REGISTER](#) form.

Once you have completed and sent the form you will have to wait until the activation of the account from the website administrator. If some more information is needed someone from our team will get in contact with you. You can find our contact data at the end of this protocol.

For those laboratories that require more than one contact per account or are in a situation not contemplated in this protocol will have to contact the organizer using the Contact tab to be instructed how to proceed.

#### **APPLICATION FOR THE PROFICIENCY TEST**

To participate in this proficiency test or to be updated via email of any changes regarding this this, it is required to have a laboratory code. To get your laboratory code is needed to apply through the website or have it arranged by the coordinator at the latest, on the deadline to request the participation.

In the “Proficiency Tests” tab on our website, you will have to select the proficiency test you want to participate, by clicking its name or the image you will enter the proficiency test page, which contains general information and a summary of that proficiency test, there you can find the present document (the protocol) and the button to start the application.

You will be required to log in to start the application and then the website will require you to submit your Limit of Quantification (LOQ) for the parameters you will study. **The analytes left as NA (NOT ANALYSED) will NOT appear in the Results form** and therefore will not be able to send results for those parameters through the website.

Once the participation has been requested, as soon as possible, it will be checked by the website administrator, and you will be sent an email with the participation code. This code will be just known only by the organizer and the participant and will be always kept confidential, even after the proficiency test finishes.

You can check on the dashboard of your client area if an application you sent has been accepted or is still pending.

Just one application per exercise can be sent by each laboratory site, not being allowed for a laboratory site to participate with two different codes.

The applications of the laboratories will be studied and accepted in base of the quantification limits of the analytes of the proficiency test and if the logistics allow the sample shipping without risk of deterioration.

According to the experience, **TestQual** can anticipate that the number of participants of this proficiency test will be around **20-30**, being 8 the minimum participants for the proficiency test to take place.

## **5. TEST MATERIAL. CONTROL AND DISTRIBUTION.**

**TestQual 211** scheme is a proficiency test based in the analysis of Multiresidue pesticides in green beans that has been spiked or were present in the matrix.

The material will be bought in a specialized shop in Spain and verified as adequate to prepare the samples by a subcontracted laboratory that holds the standard UNE-EN ISO/IEC 17025 into force.

Blank items will be reserved to be acquired along the proficiency test, the blank item can be shipped along the PT sample, saving the shipping costs. The blank items will be available at a reduced cost for the participants of this PT. Blank samples are guaranteed to have a concentration <FP/FN (concentration limit to determine False Negatives and/or False positives) for the analytes of your interest (according to the LOQ you submitted and intended analytes to be analysed).

BLANK ITEMS, if available (contact us to check), WILL BELONG TO A DIFFERENT LOT THAN THE USED TO PREPARE THE PROFICIENCY SAMPLES.

The subcontracted laboratory will analyse randomly selected sample to ensure the homogeneity and during and after the proficiency test more samples to check the stability of the parameters through the duration of the round.

Once the matrix is deemed adequate the material is cut in very small pieces or triturated, which might be assisted by dropping the matrix into liquid N<sub>2</sub>, once it is fully frozen, it is triturated. The final texture will either be a fine powder (sieved to have below 6 mm diameter), as puree or juice, always completely made of the matrix. This will be spiked with a solution with the analytes of the proficiency test and poured into a homogenizer with controlled temperature to ensure complete homogeneity.

Once the sample batch is ready, they will be stored in a temperature-controlled freezer below -20°C until the dispatch of the samples.

As previously mentioned, before the samples are distributed, the homogeneity of the lot will be assessed analysing in repeatability conditions randomly selected samples in duplicate. If the mean concentration obtained from this study is not within the planned range, the participants will be

informed, and a new distribution day might be scheduled if another spiking is deemed as necessary.

For stability assessment purpose, a total of three samples will be analysed, in duplicate, before, during and at the end (once all laboratories have sent the results) of the proficiency test.

The quality controls subcontracted by **TestQual**, including verification of adequacy of the matrix, homogeneity/stability quality controls and any other analytical study required by **TestQual** will be subcontracted to an accredited laboratory in ISO/IEC 17025 into force.

In the evaluation report will be included the conclusions and if applies, any comment regarding homogeneity &/or stability. Additionally, the results of these tests and the spiking evaluation is available to all participants of this proficiency test upon request.

The shipment of the test materials will take place on the date shown in the calendar\*, to the address provided by each laboratory in the application. The participant is responsible to communicating **TestQual** if the address is changed or update the delivery address in their client area at least 2 working days before the shipping.

\*Specific delivery dates can be changed from the scheduled dates of the calendar, but all changes will be announced either on the website or by email to the registered laboratories. **Before shipping the samples TestQual will email** the registered laboratories the **sample instructions and will confirm the definitive distribution date**, deadline to submit results and date of the report, if necessary, the dates shown in the sample instructions override the calendar here shown and the new dates will be updated in the PT page.

This PT will consist in a single round in which will be sent a sample of approximately **150 g** of test material. The samples will be sent by courier service (Nacex or DHL, depending on the destination). The material will be sent with materials and package that ensures the samples arrives correctly. The transit will be 1, 2 or 3 days to the destination country, depending on the location of the receiving laboratory.

The shipping conditions for this PT are:

FROZEN, isothermal boxes will be uses and filled with either dry ice or cold packs (-21°C) to keep the temperature.

The shipping costs are not included in the price displayed on the website, which can only be seen if you are registered and logged in. To get an approximation of the shipping costs you can get your quotation by using the contact data at the end of this protocol.

The distribution of samples will subcontract to a courier previously homologated by **TestQual**. The main criteria being the courier's delivery time to ensure the receipt of the sample is correct in the participant's facilities.

In addition to this, **TestQual** established other characteristics important for a courier like shipping management (tracking, notifications, exceptions), and ensuring the delivery conditions are proper (low breakage/lost ratio, keeping of cold chain, required documentation), always checking and evaluating they are complying with **TestQual**'s requirements.

A second test material can be requested date if necessary. If the package and/or the sample arrived damaged, defective, or not valid the participating laboratory will have to notify of this to the coordinator before two working days to get another sample.

Before the shipment, **TestQual** will send the instructions for storage and analysis via email and confirm the distribution date. You can request a paper copy to be attached to the package and/or **TestQual** might decide to include it in addition to have it sent by email.

From **TestQual** we encourage our participants to read the samples instructions carefully and follow them, as it can help to correctly conserve the sample and increase the reproducibility of the analysis. You can request a digital copy of this document (sample instructions) by letting us know through any communication channel.

## **6. CONCENTRATION RANGES, SIGMA OBJECTIVE AND ANALYTES**

For this proficiency tests the range of concentration for the target analytes might be between **10** and **400 µg/kg approximately**. This range is not absolute and depending on the various factors some or no analyte might be within this range. This range is given to provide participants a sense of the concentration ranges we usually work and expect in these kind of proficiency tests (combination of analyte/matrix).

The maximum concentration can vary depending on the MRL of the analytes in the matrix.

The **sigma objective ( $\hat{\sigma}$ )** which works in this scheme can be checked on section 9 of this protocol.



In this proficiency test, for each group of analytes, a percentage of analytes will be present (being at least >1% (the % is determined internally in our procedure) to possibly all the analytes of the list). Therefore, this PT should be useful to check each of the groups, possibly in different degrees, depending on the number of selected/present analytes and those that your laboratory analyse.

In the following table you can check the groups of analytes for this proficiency test, below you can find the lists of possible positives and how the website and the results are expected to be sent:

Analyte-group code	Grouped analytes
A	Multiresidue pesticides

The list of **possible multi-residues pesticides** in the matrix are from the list below:

2-Phenylphenol	Demeton-S-methyl sulfone	Furalaxyl	Pyrazphos
3,5-Dichloroaniline	Desmetryn	Furathiocarb	Pyridaben
3-Hydroxy-carbofuran	Dialifos	HCH-Alpha	Pyrifenoxy
4,4-Dichlorobenzophenone	Diazinon	HCH-Beta	Pirimicarb
Abamectin	Dichlofenthion	HCH-Delta	Pirimicarb-desmethyl
Acephate	Dichlormid	HCH-Gamma (lindane)	Pirimiphos-ethyl
Acetamiprid	Dichlobenil	Heptachlor	Pirimiphos-methyl
Acetochlor	Diclobutrazol	Heptachlor-epoxide	Pyriproxyfen
Aclonifen	Dichlofluanid	Heptenophos	pp-DDE
Acrinathrin	Diclofop-methyl	Hexachlorobenzene	pp-TDE(DDD)
Alachlor	Dicloran	Hexaconazole	Prochloraz
Aldicarb	Dicrotophos	Hexaflumuron	Procymidone
Aldicarb sulfone	Dieldrin	Hexazinone	Propham
Aldicarb sulfoxide	Diethofencarb	Hexythiazox	Profenofos
Aldrin	Difenoconazole	Imazalil	Profluralin
Antraquinone	Difenoxuron	Imazamethabenz-methyl	Promecarb
Atrazine	Diflubenzuron	Imidacloprid	Prometryn
Azaconazole	Diflufenican	Indoxacarb	Propachlor
Azinphos-ethyl	Dimethenamid	Iprobenfos	Propamocarb
Azinphos-methyl	Dimethoate	Iprodione	Propanil
Azoxystrobin	Dimethomorph	Iprovalicarb	Propargite
Benalaxyl	Dimoxystrobin	Isazofos	Propetamphos
Bendiocarb	Diniconazole	Isocarbophos	Propiconazole (sum of isomers)
Benfluralin	Dioxacarb	Isofenphos	Propyzamide
Benfuresate	Dioxathion	Isofenphos-methyl	Propoxur
Bentazone (Sum of bentazone, its salts and 6-hydroxy (free and	Diphenylamine	Isoproturon	Prosulfocarb

conjugated) and 8-hydroxy  
 bentazone (free and  
 conjugated), expressed  
 as bentazone)

Bifenthrin	Dipropetryn	Lambda-Cyhalothrin	Prothiofos
Bitertanol	Disulfoton	Lenacil	Pyridafenthion
Boscalid	Ditalimfos	Leptophos	Pyrimethanil
Brodifacoum	Diuron	Linuron	Quinalpho
Bromacil	Dodine	Lufenuron	Quinoxyfen
Bromocyclen	Endosulfan-alpha	Malaoxon	Quintozene
Bromophos-ethyl	Endosulfan-beta	Malathion	Rotenone
Bromophos-methyl	Endosulfan-sulfate	Mecarbam	Simazine
Bromopropylate	Endrin	Mefenpyr-diethyl	Simetryn
Bromuconazole	EPN	Mepanipirim	Spinosad A+D
Bupirimate	Epoxiconazole	Mepronil	Spirodiclofen
Buprofezin	Etaconazole	Metalaxyl	Spiromesifen
Butafenacil	Ethion	Metamitron	Spiroxamine
Butoxycarboxim	Ethoprophos	Metazachlor	Sulfotep
Butralin	Etoxazole	Methacrifos	Sulprofos
Buturon	Ethiofencarb	Methamidophos	Tebuconazole
Cadusafos	Ethiofencarb -sulfone	Methidathion	Tebufenozide
Captan	Ethiofencarb -sulfoxide	Methomyl	Tebufenpyrad
Carbaryl	Etofenprox	Methoxyfenozide	Tebupirimfos
Carbendazim	Ethofumesate	Metobromuron	Tecnazene
Carbophenothion	Etrimfos	Metolachlor	Teflubenzuron
Carbofuran	Famoxadone	Methoprotryne	Tefluthrin
Chlorantraniliprole	Famphur (Famophos)	Metoxuron	Terbacil
Chlorbromuron	Fenarimol	Metribuzin	Terbufos
Chlorfenapyr	Fenazaquin	Mevinphos	Terbumeton
Chlorfenvinphos	Fenbuconazole	Myclobutanil	Terbutylazine
Chlormephos	Fenbutaestan oxide	Molinate	Terbutryn
Chloroneb	Fenchlorphos	Monocrotophos	Tetraconazole
Chloropropylate	Fenhexamid	Monolinuron	Tetradifon
Chlorpyrifos	Fenitrothion	Monuron	Tetramethrin
Chlorpyrifos Methyl	Fenoxycarb	Napropamide	Tetrasul
Chlorthion	Fenpropathrin	Neburon	Thiabendazole
Chlorthiophos	Fenpropimorph	Nitenpyram	Thiacloprid
Cyanazine	Fenpyroximate	Nitrofen	Thiamethoxam
Cyazofamid	Fensulfotion	Nitrothal-isopropyl	Thiodicarb
Cyfluthrin (cyfluthrin including other mixtures of constituent isomers (sum of isomers))	Fenthion	Norflurazon	Thiobencarb

Cymoxanil	Phenthoate	Nuarimol	Thiometon
Cypermethrin			
(cypermethrin including			
other mixtures of	Fenuron	Ofurace	Tolclofos-methyl
constituent isomers			
(sum of isomers))			
Cyproconazole	Fenvalerate (incl. Esfenvalerate)	Omethoate	Triadimefon
Clofentezine	Fonicamid	op-TDE (DDD)	Triadimenol
Clomazone	Fluazifop-P-butyl	Oxadiazon	Triazophos
Cloquintocet-mexyl	Fluchloralin	Oxadixyl	Trichloronate
Chlorfenson	Flucythrinate	Oxamyl	Tridemorph
Chlorotoluron	Fludioxinil	Oxamyl-oxime	Trifloxystrobin
Chloroxuron	Flufenoxuron	Oxydemeton-methyl	Triflumuron
Chlorpropham	Flumetralin	Oxyfluorfen	Trifluralin
Chlorsulfuron	Fluometuron	Paclobutrazol	Vinclozolin
Chlorthal-dimethyl	Fluotrimazole	Parathion	Yodofenfos
Clothianidin	Fluquinconazole	Parathion-methyl	Zoxamide
Coumaphos	Flusilazole	Pebulate	
Kresoxim-methyl	Flutolanil	Penconazole	
Crimidine	Flutriafol	Pendimethalin	
Cyanofenphos	Folpet (phthalimide)	Pentachloroanisole	
Cyanophos	Fonofos	Permethrin	
Cycloxydim	Formothion	Phenmedipham	
Cyprodinil	Phosalone	Picoxystrobin	
Deltamethrin	Phosphamidon	Piperonyl butoxide	
Demeton-S-methyl	Phosmet	Pyraclostrobin	

## 7. RESULTS EXPRESSION

Each participant laboratory must analyse the sample received according to their routine procedure and fill up the RESULTS form of its client are of the website [www.TestQual.com](http://www.TestQual.com) with just one value per analyte/parameter.

The results should be expressed in **µg/kg**. The number of significant figures and the units are to be chosen by laboratories and will be displayed in the report as received by the participant.

The method used for the analysis of each compound informed should be sent when filling up the results form.

The organizer should get the results before the fixed data of the scheme.

If you have any problem logging in to your client area or submitting your results you can contact the coordinator of the PT for guidance or help.

Once the results are sent you can check if they were correctly recorded by accessing the detailed information of this proficiency test, which can be accessed in your client area. You are allowed to modify results within the stipulated period range, until the deadline to submit results, included. Currently our website, once the results are sent, cannot be changed, therefore, for any modification you will be required to open an issue in your client area for this PT and send the required modifications through that communication channel.

## 8. STATISTICAL EVALUATION

TestQual will develop the following statistical evaluation:

TestQual considers as an **extreme outlier** any data which differs more than **50 %** of the average of all results reported by the laboratories, according to the Harmonize Protocol of the IUPAC. These extreme values will not be included in the calculation of the assigned value.

Once received all the results, **TestQual** evaluates the unimodality of all the valid values with a test using the kernel's density representation, being explained in the final report which is the followed procedure in case there is more than one distribution.

If enough valid results (determined in our internal procedure) are provided and the uncertainty is negligible the **assigned value (X)** will be determined using the robust average of the results considered valid for statistical computing (after eliminating the extreme outliers), according to the standard ISO 13528 into force.

In case of having a lower number of results than expected, the participants would be evaluated against an independent value from their own results, since the lower the number of results is used to obtain the assigned value, the higher is the chance of not identify necessary action signals in case the participants have/had an unknown bias. For this reason, both the assigned value obtained from robust statistic and the characterization value are compared between them, this way we check that there are not sustainable differences and depending on the number of results available and the uncertainty of the assigned value, TestQual will chose according to our internal procedure how the assigned value will be determined and evaluate participants, accordingly, showing it in the report.

If the consensus value is not used, then the characterization of the spiking and the uncertainty will be calculated based in a suitable model and according to ISO 13528 into force.

Likewise, and as mentioned, the characterization value will be compared to the participants' consensus and only used if there are not sustainable differences between assigned values. If the assigned values are too different, no evaluation or informative evaluation would be issued, explaining the reason in the report.

The **standard uncertainty ( $u_x$ )** when the assigned value is obtained from the consensus it is calculated using robust statistics from the following formula:

The **standard uncertainty ( $u_x$ )** is calculated using robust statistics from the following formula:

$$u_x = s^*/\sqrt{p}$$

Being  $s^*$  the robust standard deviation of the data and  $p$  the number of results considered.

When the assigned value is obtained from the spiking done, the uncertainty is calculated accounting all sources of uncertainty according to our internal procedure and according to ISO 13528 into force.

The following condition must be fulfilled to discard the contribution of the uncertainty:

$$u_x \leq 0,3 \hat{\sigma}$$

The **standard deviation for proficiency assessment**, also named **target standard deviation, ( $\hat{\sigma}$ )**, comes from this formula:

$$\hat{\sigma} = b_i \cdot X$$

Being  $b_i = \%_{tRSD} / 100$ , and  $\%_{tRSD}$  is the target relative standard deviation.

In this case, the assigned relative standard deviation is **25 %**.

This value has been chosen according to the experience of the organizer in this and similar proficiency test which procedure of the analysis like the parameters of this PT.

**Proficiency assessment (z-score):** This parameter shows the competence and accuracy of the laboratory. It is calculated using the following formula:

$$z = (x_i - X) / \hat{\sigma}$$

Where  $x_i$  is the value reported by the laboratories,  $X$  is the assigned value, and  $\hat{\sigma}$  is the target standard deviation for each analyte.

The criteria for defining the z-score values are:

	$ z  \leq 2$	Satisfactory
$2 <$	$ z  \leq 3$	Questionable
	$ z  > 3$	Unsatisfactory

In case the inequation  $u_x \leq 0,3 \hat{\sigma}$  is not fulfilled, the participants of the scheme will be informed in the report that the uncertainty is not negligible. For the parameters/analytes in which this situation occurs, the following calculation will be made:

$$z'\text{-score} = (x_i - X) / \sqrt{\hat{\sigma}^2 + U_x^2}$$

Where  $x_i$  is the value reported by the laboratories,  $X$  is the assigned value,  $\hat{\sigma}$  is the target standard deviation for each analyte and  $U_x$  is the uncertainty of the assigned value.

The criteria for defining the z'-score values are:

	$ z'  \leq 2$	Satisfactory
$2 <$	$ z'  \leq 3$	Questionable
	$ z'  > 3$	Unsatisfactory

The z'-score is a underestimation of the z-score, for this reason, for those analytes in which the uncertainty of the assigned value cannot be neglected and a z'-score is issued, it will be accompanied by the percentual difference against z-score, this way participants should be able to complete evaluate their performance.

The evaluation could be informative if the difference between scores surpasses the limit contemplated in our procedure. If any analyte or evaluation is informative it will be indicated in the report through marking and a legend.

Both z-score and z'-score will be issued accredited if the criteria of our technical annex are met.

False negatives: Any analyte not reported in the results that is in the sample above the limit of quantification previously established for this proficiency test by the organization and above the LOQ of the participant laboratory (**10 µg/kg**). **TestQual** assigns to all false negatives a result equal to half the laboratory limit of quantitation (LOQ/2). If no LOQ is available, the result evaluated will be 0. If <LOQ was reported, and the assigned value is above the LOQ a false negative will be evaluated.

False positives: Those analytes reported in the results, which is not present in the test material, and is reported by the participant at concentrations higher than the limit of quantification of the P.T. (**10 µg/kg**).

#### **Testing for sufficient homogeneity:**

Ten samples will be chosen at random and sent to be analysed by **TestQual**'s subcontracted laboratory in duplicate in repeatability conditions.

Once received the results, a statistical evaluation will be performed, with the homogeneity evaluation of the Harmonic Protocol published by IUPAC and based in ISO 13528 into force.

The acceptance criterion to ensure that the randomly chosen samples are homogeneous is that the square of the estimated sampling standard deviation is below the critical value for accepting proper homogeneity:

$$S_{\text{sam}}^2 < c$$

Being  $S_{\text{sam}}^2$  the estimated sampling standard deviation, obtained from the variance of the results sums and the experimental estimate of analytical standard deviation ( $S_{\text{an}}$ ), which in turn is obtained from the differences between replicates of the same sample. Lastly,  $c$  is the limit value, which is obtained as ISO 13528 states, it considers two constants, obtained from two significance groups, which are multiplied with a term related to the target standard deviation and the  $S_{\text{an}}$ .

If  $S_{\text{sam}}^2 < c$  is true, then the lot prepared will be considered sufficiently homogeneous and only then it would be distributed. With this test what we achieve is to check if the intra-sample deviation is lower than the inter-sample deviation.

If the results from the homogeneity test does not meet the criteria **TestQual** would communicate any change required in the proficiency test (new lot will be prepared, new calendar, etc.).

**Testing for sufficient stability:**

Three samples will be analysed, in duplicate, before, during and at the end (once all laboratories have sent the results) of the proficiency test.

With these values, a study is performed according to the up-to-date SANTE guide (*SANTE/11312/2021 Guidance document on analytical quality control*), referred to analysis under repeatability conditions.

The acceptance criteria to ensure the samples have been stable during the whole proficiency test are the following:

$$| (X_{t1} - X_{t2}) / X_{t1} | \cdot 100 \leq 10\%$$

$$| (X_{t1} - X_{t3}) / X_{t1} | \cdot 100 \leq 10\%$$

Being  $| (X_{t1} - X_{tn}) / X_{t1} |$  the difference between the average of the samples analysed before, during and at the end of the proficiency test.

If this criterion is not met, **TestQual** will review if the instability can be neglected using statistical, data driven and/or technical approaches.

**9. EVALUATION REPORT/S**

Upon completion of the analysis period **TestQual** will conduct a comprehensive statistical evaluation of all participant laboratories' results. Subsequently, **TestQual** will either issue an individual report, summarizing each participant's performance, or a global report **according to the published calendar**. Regardless of the type of report, the reports will be sent via email in PDF format or made available for download from the client area on [www.TestQual.com](http://www.TestQual.com).

**Either** if a global report is issued or an individual report is issued, participants are given a 30-day period to review and, if necessary, lodge an appeal against the evaluation of the proficiency program or report any identified mistakes **from TestQual side**. Following this appeal window, **TestQual** will proceed to release the final global report within the **upcoming 30** days, with the complete performance of all laboratories **and appellations accepted (detailed in report if accepted)**.

This allows participants to communicate concerns or disputes related to results, evaluation received, or any mistakes identified **from TestQual side**. Reports can be appealed for mistakes



found even outside the formal appellation period, ensuring a mechanism for ongoing quality assurance.

Participants have the option of requesting analyte-group reports, specific to the groups of analytes given in the proficiency test. It's important to note that these individual reports are not accredited and serve as a summary, cross-referencing the global report to ensure comprehensive information is given. All participants that send results will receive at least the global report.

This structured process ensures fast reporting, transparency, accountability, and the provision of comprehensive reports tailored to the specific needs and interests of each participant in **TestQual's** proficiency testing program.

## 10. CONTACT

**TestQual** puts at your disposal any of the following means to contact our team:

Website:	<a href="http://www.TestQual.com">www.TestQual.com</a>
Email:	<a href="mailto:testqual@testqual.com">testqual@testqual.com</a>
Office phone:	+34 868 94 94 86
Mobile phone:	+34 676 367 555

## 11. BIBLIOGRAPHY AND REFERENCES OF INTEREST

**TestQual** Proficiency Testing Schemes are based on the following standards:

UNE-EN ISO/IEC 17043, into force. Conformity assessment- General requirements for proficiency testing.

ISO13528 into force. Statistical methods for use in proficiency testing by interlaboratory comparison.

THE INTERNATIONAL HARMONIZED PROTOCOL FOR THE PROFICIENCY TESTING OF ANALYTICAL CHEMISTRY LABORATORIES

EU Pesticides database (v.2.2) Search Pesticides residues

SANTE/11312/2021 v2 (Implemented by 01/01/2024) Guidance document on analytical quality control and method validation procedures for pesticides residues analysis in food and feed.

**END OF DOCUMENT**

